



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES  
STATE PUBLIC HEALTH LABORATORY  
**ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT**

**RECEIVED**

By Carol Day at 6:39 am, Feb 09, 2015

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 030450	PRINTER SN 95.1111.053	DATE OF INSPECTION 02-07-2015
LOCATION OF INSTRUMENT (STREET AND CITY) 501 Faraon St. Joseph MO 64501		TIME OF INSPECTION 2106

**CHECKLIST:** Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

☒ DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)

☒ TEMPERATURE OF ALCO SENSOR (1 O-C - 40-C)

☒ PRINTER WORKING PROPERLY

☒ TIME AND DATE DISPLAYING PROPERLY

**BREATH ALCOHOL ACCURACY STANDARDS**

☐ SIMULATOR SOLUTION

☒ COMPRESSED ETHANOL-GAS MIXTURE

☒ STANDARD SUPPLIER Intoximeters LOT # AG428002 EXP. DATE 10-07-2016

☐ SIMULATOR TEMPERATURE (34°C ± 0.2°C) \_\_\_\_\_ SIMULATOR SN \_\_\_\_\_ SIMULATOR EXP DATE \_\_\_\_\_

☒ CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

☒ 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE

☐ 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE

☐ 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 - .100	TEST 2 - .099	TEST 3 - .099
------------------	------------------	------------------

☒ RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:  
(DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	0	(0-.04)	0	(.05-.09)	6	(.10-.14)	6	(.15-.19)	2	(OVER .19)	2
----------	---	---------	---	-----------	---	-----------	---	-----------	---	------------	---

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

Recalibrated instrument to a .100 compressed ethanol-gas mixture.

**INSPECTING OFFICER**

SIGNATURE 	PRINT NAME Wayne Byrom
TYPE II PERMIT NUMBER/EXPIRATION DATE 240241 05-19-2016	TELEPHONE NUMBER 816-271-5359

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office  
2875 James Boulevard  
Poplar Bluff, MO 63901



Airgas USA LLC (LAB)  
3500 Bernard Street  
St. Louis, Mo. 63103  
Ph: (314) 533-3100  
Fax: (314) 533-7328

## Certificate of Analysis

Customer Name  
Intoximeters, Inc.  
2081 Craig Road  
St. Louis, Mo 63146

Test Date: 8-Oct-2014

Lot # AG428002

<u>Exp. Date</u>	<u>Cyl. Type</u>	<u>Component</u>	<u>Certified Concentration</u>
7-Oct-2016	108	Ethanol	0.100 $\pm$ 2% BrAC (272 ppm)
		Nitrogen	Balance

Certification Traceable to N.I.S.T. RGM Ethanol Standards:

<u>Serial No.</u>	<u>Concentration</u>	<u>Serial No.</u>	<u>Concentration</u>
EB0010581	391.8 ppm	EB0010603	392.5 ppm
EB0010570	259.8 ppm	EB0010559	258.9 ppm
EB0010285	209.0 ppm	EB0010595	208.9 ppm
EB0010561	103.7 ppm	EB0010562	104.9 ppm
EB0010681	52.22 ppm	EB0010579	52.94 ppm

Analytical Method: NDIR

Digitally signed by Quality Control  
Date: 2014.10.08 12:15:00 -05:00  
Reason: Dry gas standard certification of analysis  
Location: Airgas USA LLC (Lab)

Analyst: \_\_\_\_\_

Rod Marsala

ISO 17025:2005 A2LA accredited. Certificate Number 2989.01

AS IV Serial no: 000450  
Version no: 7410

TEST RECORD 05954

Temp Date Time 210L  
s/  
Air Blank:  
02/07/15 21:06 .000  
Calibration Check:  
25 02/07/15 21:06 .100

Subject Name

Subject I.D.

Operator Name, I.D.

Location

AS IV Serial no: 000450  
Version no: 7410

TEST RECORD 05955

Temp Date Time 210L  
s/  
Air Blank:  
02/07/15 21:09 .000  
Calibration Check:  
25 02/07/15 21:09 .099

Subject Name

Subject I.D.

Operator Name, I.D.

Location

AS IV Serial no: 000450  
Version no: 7410

TEST RECORD 05956

Temp Date Time 210L  
s/  
Air Blank:  
02/07/15 21:11 .000  
Calibration Check:  
25 02/07/15 21:11 .099

Subject Name

Subject I.D.

Operator Name, I.D.

Location

AS IV Serial no: 000450  
Version no: 7410

TEST RECORD 05957

Temp Date Time 210L  
s/  
Air Blank:  
02/07/15 21:15  
Void: RPI  
12 02/07/15 21:15

Subject Name

Subject I.D.

Operator Name, I.D.

Location



STATE OF MISSOURI  
DEPARTMENT OF HEALTH AND SENIOR SERVICES  
BREATH ALCOHOL PROGRAM

2

**PERMIT**  
**TYPE II**  
**WAYNE BRYOM**

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

**ALCO-SENSOR IV WITH PRINTER, DATAMASTER**

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 5/19/2014

NUMBER 240241

EXPIRES 5/19/2016

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 680-0771 (6-10)

LAB-4 (16-10)

STATE OF MISSOURI  
DEPARTMENT OF HEALTH AND SENIOR SERVICES  
BREATH ALCOHOL PROGRAM

**INSTRUMENT OPERATOR CARD**

*The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.*

Operator BRYOM, WAYNE  
Permit No 240241  
Date Issued 5/19/2014 Date Expires 5/19/2016